

Test #28 Chamber ME1.2.003

2003-4-24 13:41

HV, V	ALCT, kHz	CLCT, k Hz
3100	0.001	0.000
3200	0.006	0.000
3250	0.018	0.001
3300	0.042	0.003
3350	0.073	0.007
3400	0.095	0.020
3450	0.102	0.026
3500	0.108	0.033
3550	0.109	0.033
3600	0.109	0.032
3650	0.109	0.033
3700	0.109	0.033
3750	0.111	0.036
3800	0.111	0.038

